



CERTIFICATE

No. Z1 04 11 39096 006

Holder of Certificate: **Völker AG**
Wullener Feld 79
58454 Witten
GERMANY

Certification Mark:



Product: **Care Beds**
Electrically adjustable care beds

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. See also notes overleaf.

Test report no.: 70081417



Date, 2004-12-03

Page 1 of 2



CERTIFICATE
No. Z1 04 11 39096 006

Model(s): **Electrically adjustable care bed**
Type: 3080

Parameters:

total length:	2070 mm
total width:	980 mm
mattress support platform:	2000 mm x 900 mm, four sections
controller:	Okimat 480, co. Okin
motor stroke rack:	2x OG 01, co. Okin
motor backpart:	Okimat 480, co. Okin
motor footpart:	Okimat 480, co. Okin
accumulator:	accumulator controller, AK-XX-1, 2 Ah, co. Okin (optional)
locking device:	switch box with 8 buttons and rotary switch, IP X4, co. Okin
manual switch:	special design with 7 buttons, co. Völker, distance between buttons 9 mm
class of protection:	II
degree of protection:	IP X4
transformer rated voltage:	U _{in} : 230 V AC, 50/60 Hz, 150 W
cyclic duration factor:	15% (max. continuous operation: 2 minutes, max. switch cycles: 5 per minute)
retractable from:	410 mm to 801 mm
undercarriage UG 35:	the possibility for going under is limited on 130 mm
max. patient weight:	175 kg
max. working load (SWL):	210 kg

Remark:
 The section 52.5.102 of the DIN EN 60601-2-38: 2001 can be fulfilled only with the optional accumulator.

Tested according to: EN 60601-2-38/A1:2000
 EN 1970:2000
 (in Anlehnung / with regard to)

Production Facility(ies): 39096